

Study Type- Original Research

COMPARATIVE EVALUATION OF INHERENT ANTIMICROBIAL PROPERTIES AND BACTERIAL SURFACE ADHERENCE BETWEEN COPPER AND STAINLESS STEEL SUCTION TUBE

Running title – Anti microbial properties of copper vs stainless steel suction tube.

Abstract

Aim of the present study is comparative evaluation of the inherent antimicrobial properties of the copper and stainless steel suction tube also evaluate bacterial adherence and reduction with time within the suction tube. Present study was performed in an in vitro setting in Saveetha dental college and hospital. Both the metals suction tubes were cut into equal lengths of 2.3 inches and dipped in a sterile container for 5 minutes which had bacterial suspension of *Streptococcus Mutans*. Each group included three samples of copper and stainless steel suction tubes. Control group was taken out at 0 mins post initial dipping period de-coated in a sterile container and immediately cultured on Brain heart infusion agar, followed by group 1 at 15 minutes group 2 at 60 minutes and group 3 at 180 minutes respectively. Bacterial colonies were counted after overnight incubation. Copper suction tubes demonstrated less initial adhesion of bacterial colonies than that of stainless steel suction tubes. Percentage reduction in bacterial colonies was observed more in stainless steel suction tubes with respect to time. Spite of high initial adhesion stainless steel suction tube demonstrated higher degree of percentage reduction in bacterial colonies with time hence proven better for antimicrobial properties than copper suction tube.

Key words: antimicrobial property, cross infection, copper, contact killing, stainless steel, suction tube, *S.Mutans*

1.Introduction :

When certain bacteria come in contact with the uncoated copper surface it results in formation of intracellular oxidative stress in the bacterial cell wall due to release of ions from the copper surface which results in bacterial cell lysis^{1,2}. This phenomenon was well known since ancient times but, recently it has gotten renewed attention from the researchers. The term ‘contact killing’ was coined for the above phenomenon. United States Environmental Protection Agency (US EPA) recognized copper as the first antimicrobial metal in the year of 2008³. One of the most important advantages of copper as an antimicrobial agent is its low levels of resistance in the microorganisms³.

Healthcare associated infection(HAI) stands for or iatrogenic infections are one of the most important causes of increased hospitalization and complications post procedures⁴⁻⁶. Factors playing a role in the spread of these infections include patient condition, hospital or dental clinic

environment and mainly contaminated surfaces and instruments pre procedure and even during the procedure due to improper sterilization protocol^{2,7-11}. Therefore meticulous sterilization protocol and surface disinfection is of utmost importance to avoid iatrogenic propagation of infection and postoperative complications¹². In recent years coating the surfaces with solid copper or mixing copper alloys has been practiced in hospitals to reduce the number of HIA's. While some studies show promising results or reduction in infections others find the use of copper to be non significant. As most of the studies have highly variable study design and setting it is hard to compare different studies although use of copper in medical devices and workplaces shows a promising future^{3,13-15}.

Removal of the body fluids and irrigants (eg. Sodium hypochlorite, saline, saliva etc.) is of high importance in modern dental as well as medical practice. Potent evacuation of the fluids aids in improving the quality of the treatment also in comfort of the operator as well as the patient. The suction unit becomes contaminated by residual microbes and is considered the most common cause of the HAI's.

Suction in the dental chair works on the vacuum⁴⁻⁶. This unit is repeatedly used only with a change of the suction tip for consecutive patients. Such practice although not considered safe and risk free by the researchers is yet followed all over the world^{3,13-15}. Disposable plastic suction tubes are available in pre sterilized packs but to reduce the risk during the major surgical procedures (autoclaved) metal suction tubes are preferred over the pre-sterile plastic suction tubes as they can be sterilized before each patient and has least chances of contamination unlike pre-sterile plastic suction tube.³

Chances of cross contamination although low still hold a threat for the major surgical procedures and can cause postoperative complications³. First case of cross contamination due to back flow of fluids from the suction tube was recorded in the year of 1993 and a recent one in the year of 2006¹⁶.

Hence in this study we decided to compare suction tubes made with solid copper against the conventional stainless steel suction tube to evaluate whether the intrinsic antimicrobial property of the copper has any significance in the reduction in the risk of the procedures¹⁷⁻¹⁹.

These are some of the previously undertaken research topics in the field of periodontology by Saveetha Dental College²⁰⁻³². Now we are planning to do a comparative evaluation of intrinsic antibacterial activity of copper suction tube vs stainless steel suction tube against the *S.mutans* strains.

2. Material And Methods:

Present study is an in vitro study carried out in saveetha dental college and hospital chennai. *Streptococcus Mutans* (figure 3) was selected to evaluate the antimicrobial properties of both the metal suction tubes (figure 1). Study was carried out under expert supervision of a guide in the microbiology department of Saveetha dental college and hospital. Study was done after acquiring all necessary clearances from the college review board and research department. Both copper and stainless steel suction tubes were sectioned into 2.3 inches pieces so that they fit in the selected sterile airtight containers. The tubes were autoclaved separately and divided into 4 groups (group 0(0 mins) , group 1(15mins), group 2(60 mins) , group 3(180 mins). Each group included 3 tubes of copper and stainless steel. Bacterial suspension of *S. mutans* was made in a sterile airtight container and all the tubes were completely immersed in the suspension for 5 minutes. Post emersion tubes were segregated into groups according to time interval mentioned above. Control group tubes were immediately de-coated in separate Eppendorf tubes and 0.2ml lawned using the zig-zag strokes (figure 4) on brain heart infusion agar culture media. Similar protocol was followed for the remaining groups. Group 1 group 2 and group 3 were de-coated and cultured at 15 minutes, 60 minutes and 180 minutes respectively on the same media (separate petri plates) and incubated overnight in an automated incubator. Bacterial colonies were counted using Scan 300 - Colony counter for each group post 24 hours of incubation by the primary examiner and recounted by a third person who was not part of the study and average was taken to calculate the results to minimize the assessment bias.

3.Results And Discussion :

Result table :

Table 1: The number of *S.Mutans* colonies at different time intervals in copper and stainless steel suction tubes.

Metal /Time	Control		Group 1 (15 mins.)		Group 2 (60 mins.)		Group 3 (180 mins.)	Mean
Copper	730	738	140	120	40	32	17	17
	772		112		22		13	
	700		128		34		22	
Stainless steel	1000	1189	127	137	37	47	10	18
	1378		147		57		18	
	1189		137		47		28	

*Mins = minutes

Table 2: The percentage of the remaining number of bacterial colonies with respect to time seen in each time interval in copper and stainless steel suction tubes.

	Percentage of remaining colonies after each time interval		
Metal/Time	From control to group 1	From group 1 to group 2	From group 2 to group 3
Copper suction tube	16.26%	4.33%	2.30%
Stainless steel suction tube	11.52%	3.98%	1.51%

Table 3: The percentage reduction in the number of bacterial colonies (from bacterial colony count at 0 mins.) with respect to time seen in each time interval in copper and stainless steel suction tubes.

	Percentage of reduction in the total number of bacterial colonies after each time interval		
Metal/Time	From control to group 1 (15 mins)	From group 1 to group 2 (60 mins)	From group 2 to group 3 (180 mins)
Copper suction tube	83.74%	95.67%	97.70%
Stainless steel suction tube	88.48%	96.02%	98.49%

*Mins = minutes



Figure 1 : Stainless steel and copper suction tubes in separate autoclaved.



Figure 2: Sterile airtight holding containers with labels according to the groups.



Figure 3: *Streptococcus Mutans* strain used for the study.




Figure 4: Process of lawn culturing on the brain heart infusion agar plate.




Figure 5: Bar diagram showing number of bacterial colonies at each time interval. Blue colour bar represents values of copper group and green colour bar represents the values of stainless steel group. Where x axis represents the time intervals and colour bars represent the two different material groups included in the study. As we can see at 0 mins the number of colonies in the copper group are significantly less than that of the SS group which implies less initial adhesion of bacteria to Cu suction tubes than that of copper. (738 in Cu and 1189 inn SS respectively.) We can also see that the percentage reduction in the SS group at each time interval is higher than that of the Cu group. T 180 mins both the groups show 17 and 18 number of colonies in Cu and SS groups respectively.

At 0 mins the number of colonies in the copper group are significantly less than that of the SS group which implies less initial adhesion of bacteria to Cu suction tubes than that of copper. (738 colonies in Cu and 1189 colonies in SS respectively.) We can also see that the percentage reduction in the SS group at each time interval is higher than that of the Cu group. T 180 mins both the groups show 17 and 18 number of colonies in Cu and SS groups respectively.

As seen in the table 1 (figure 5) initial bacterial adhesion on the stainless steel suction tube was significantly higher than that of the copper suction tubes at 0 minutes (control group) which was 738 colonies for copper suction tube and 1189 colonies for stainless steel suction tubes respectively. Group 1 which was cultured at 15 mins showed 120 and 137 colonies in copper and stainless steel suction tubes respectively. Group 2 which was cultured at 60 mins showed 32 and 47 colonies in copper and stainless steel suction tubes followed by group 3 which was cultured at 180 mins and showed 17 and 18 colonies in copper and stainless steel suction tubes respectively.

In the present study we observed that the initial adhesion on stainless steel tubes was significantly higher and the time interval in the number of colonies was much higher than that of the copper suction tubes as well this implies that stainless steel suction tube showed better antimicrobial effect on *S.Mutans*. Within the limits of the study it can be concluded from the results in table 2 that the drop in the number of colonies in case of stainless steel suction tubes was higher than that of the copper suction tubes at 0 min to 15 mins interval which was 88.48% drop in stainless steel suction and 83.74% for copper suction tubes respectively.

Suction lines in the sterile dental clinical areas should be cleaned every day and flushed using antiseptic solution after every patient. Even after flushing the dental suction unit meticulously after each patient with non foaming antiseptic solution there is evidence that shows some amount of residual bioburden still remains within the suction unit. Changing the suction tip for every patient is mandatory¹². To further minimize the contamination of the surgical site due to the dental suction unit and the backflow, newer material with intrinsic antimicrobial properties should be promoted such as copper. A study shows that type of suction used is equally important. For example high vacuum suction shows less residual bioburden and less number of viable bacterial mass after regular flushing with the antiseptic solution than that of the regular suction^{16,17}.

It has been proven that frequently touched surfaces in the dental clinics and hospitals act as reservoirs of the residual microbial flora and are the main causes of the HAI's and postoperative complications^{2,33-35}. Multiple studies have been done in the medical field which shows a great level of evidence that the coated surfaces with metals containing intrinsic antimicrobial activity result in reduced microbial burden in the hospitals and reduce the chances of postoperative as well as intra operative complications^{2,36-39}. Covering the bed handles, bedside table surfaces, door handles with copper has helped to significantly to reduce the bacterial load⁴⁰. Very few studies have been carried out in dental practice to incorporate such metals used for instruments used on a daily basis to reduce the bacterial load in the environment, chances of cross contamination during procedures and post procedures. Contact killing property of copper metal was discovered in the early year of 2008 although it has been used widely since ancient times for its other beneficial mineral contents. Prior to discovery of antibiotic agents in 1930's copper iodine and other substances with antimicrobial properties were the epicenter of the research^{3,13-15}. Post discovery of antibiotics research about such elements was hampered but evidence of multi drug resistant bacteria's in 1980-1990 forced the research community to drive the research back to the naturally available antimicrobial elements such as copper which showed low levels of bacterial resistance³⁶.

Some of the authors found highly significant results for the intrinsic antimicrobial properties and its merits in the medical field whereas some did not^{2,7-11}. Due to variability in the study designs and lack of research in the dental field regarding use of copper incorporated materials it is difficult to formulate any hypotheses which have been proven. More research in the field is necessary to find better ways of minimizing the chances of cross contamination and postoperative complications due to iatrogenic infections^{3,13-15}. Multiple factors play a role in the outcome of any study such as surface texture of the internal surface of the suction tube, operator bias etc. which can hamper the results.

Most of the regular dental procedures are finished within 40-60 minutes hence reducing the bacterial load in the suction tube within this period is of high importance to avoid infection and cross contamination^{3,13-15}. As we can see in the results of this study (table1) copper suction tube shows fairly less number of initial bacterial adhesion, we can imply that copper suction tubes can be used for short procedures. Further research using different strains of oral pathogenic bacteria has to be done to prove copper suction tubes as a better substitute for stainless steel as well as plastic suction tubes^{2,7-11}. It will also aid in reducing the amount of dental non-degradable waste of plastic suction tubes created.

4. Conclusion:

In Spite of lesser initial adhesion on the copper surface at 0 minutes which was 738 colonies and that for SS group was 1189 colonies the percentage reduction seen on the surface of stainless steel suction tubes was higher than that of the copper suction tubes at each time interval as seen in the results (figure 5). Further research using different strains of oral pathogenic bacteria has to be done to prove copper suction tubes as a better substitute for stainless steel as well as plastic suction tubes.

References:

1. Schmidt MG, von Dessauer B, Benavente C, et al. Copper surfaces are associated with significantly lower concentrations of bacteria on selected surfaces within a pediatric intensive care unit. *Am J Infect Control* 2016; 44: 203–209.
2. Vincent M, Duval RE, Hartemann P, et al. Contact killing and antimicrobial properties of copper. *Journal of Applied Microbiology* 2018; 124: 1032–1046.
3. Montero DA, Arellano C, Pardo M, et al. Antimicrobial properties of a novel copper-based composite coating with potential for use in healthcare facilities. *Antimicrob Resist Infect Control* 2019; 8: 3.
4. Allegranzi B, BagheriNejad S, Combescure C, et al. Burden of endemic health-care-associated infection in developing countries: systematic review and meta-analysis. *Lancet* 2011; 377: 228–241.

5. Barahona N, Rodriguez M, De Moya Y. IMPORTANCIA DE LA VIGILANCIA EPIDEMIOLÓGICA EN EL CONTROL DE LAS INFECCIONES ASOCIADAS A LA ATENCIÓN EN SALUD. *Biociencias* 2019; 14: 79–96.
6. Barriga J, Cerda J, Abarca K, et al. Infecciones asociadas a la atención en salud (IAAS) en pacientes pediátricos post-operados de cardiopatías congénitas. *Revista chilena de,* https://scielo.conicyt.cl/scielo.php?pid=S0716-10182014000100002&script=sci_arttext&tlng=e (2014).
7. Brenner P, Nercelles P, Pohlenz M, et al. Costo de las infecciones intrahospitalarias en hospitales chilenos de alta y mediana complejidad. *Revista chilena de infectología* 2003; 20: 285–290.
8. Attaway HH 3rd, Fairey S, Steed LL, et al. Intrinsic bacterial burden associated with intensive care unit hospital beds: effects of disinfection on population recovery and mitigation of potential infection risk. *Am J Infect Control* 2012; 40: 907–912.
9. Kramer A, Schwebke I, Kampf G. How long do nosocomial pathogens persist on inanimate surfaces? A systematic review. *BMC Infect Dis* 2006; 6: 130.
10. Rutala WA, Weber DJ. Guideline for disinfection and sterilization in healthcare facilities, 2008, <https://stacks.cdc.gov/view/cdc/47378> (2008).
11. Prado V, Vidal R, Durán C. Aplicación de la capacidad bactericida del cobre en la práctica médica. *Revista médica de Chile* 2012; 140: 1325–1332.
12. Boyle MA, O'Donnell MJ, Russell RJ, et al. Overcoming the problem of residual microbial contamination in dental suction units left by conventional disinfection using novel single component suction handpieces in combination with automated flood disinfection. *J Dent* 2015; 43: 1268–1279.
13. von Dessauer B, Navarrete MS, Benadof D, et al. Potential effectiveness of copper surfaces in reducing health care-associated infection rates in a pediatric intensive and intermediate care unit: A nonrandomized controlled trial. *Am J Infect Control* 2016; 44: e133–e139.
14. Sifri CD, Burke GH, Enfield KB. Reduced health care-associated infections in an acute care community hospital using a combination of self-disinfecting copper-impregnated composite hard surfaces and linens. *Am J Infect Control* 2016; 44: 1565–1571.
15. Marais F, Mehtar S, Chalkley L. Antimicrobial efficacy of copper touch surfaces in reducing environmental bioburden in a South African community healthcare facility. *J Hosp Infect* 2010; 74: 80–82.
16. Kalsi A, Balani N. Cleaning, disinfection and sterilisation. *Physics for the Anaesthetic Viva*; 103–106.
17. Kenley RS. *Devices, systems, and methods for cleaning, disinfecting, rinsing, and priming blood separation devices and associated fluid lines.* 8562908,

<https://patentimages.storage.googleapis.com/a2/35/50/ba25c0b6775fc7/US8562908.pdf>
(2013, accessed 4 July 2020).

18. Robertshaw RG. Automatic cleaning and disinfection of suction bottles. *Journal of Hospital Infection* 1982; 3: 299–302.
19. De R. Cleaning, Sterilisation and Disinfection. *Diagnostic Microbiology (For DMLT Students)* 2007; 12–12.
20. Ezhilarasan D, Apoorva VS, Ashok Vardhan N. Syzygiumcumini extract induced reactive oxygen species-mediated apoptosis in human oral squamous carcinoma cells. *J Oral Pathol Med* 2019; 48: 115–121.
21. Gajendran PL, Parthasarathy H, Tadepalli A. Comparative evaluation of cathepsin K levels in gingival crevicular fluid among smoking and nonsmoking patients with chronic periodontitis. *Indian J Dent Res* 2018; 29: 588–593.
22. Kaarthikeyan G, Jayakumar ND, Sivakumar D. Comparative Evaluation of Bone Formation between PRF and Blood Clot Alone as the Sole Sinus-Filling Material in Maxillary Sinus Augmentation with the Implant as a Tent Pole: A Randomized Split-Mouth Study. *J Long Term Eff Med Implants* 2019; 29: 105–111.
23. Arjunkumar R. Nanomaterials for the Management of Periodontal Diseases. In: Chaughule RS (ed) *Dental Applications of Nanotechnology*. Cham: Springer International Publishing, pp. 203–215.
24. Ravi S, Malaiappan S, Varghese S, et al. Additive Effect of Plasma Rich in Growth Factors With Guided Tissue Regeneration in Treatment of Intrabony Defects in Patients With Chronic Periodontitis: A Split-Mouth Randomized Controlled Clinical Trial. *J Periodontol* 2017; 88: 839–845.
25. Kavarthapu A, Malaiappan S. Comparative evaluation of demineralized bone matrix and type II collagen membrane versus eggshell powder as a graft material and membrane in rat model. *Indian J Dent Res* 2019; 30: 877–880.
26. Murthykumar K, Arjunkumar R, Jayaseelan VP. Association of vitamin D receptor gene polymorphism (rs10735810) and chronic periodontitis. *J Invest Clin Dent* 2019; 10: e12440.
27. Ramesh A, Vellayappan R, Ravi S, et al. Esthetic lip repositioning: A cosmetic approach for correction of gummy smile - A case series. *J Indian Soc Periodontol* 2019; 23: 290–294.
28. Ramesh A, Varghese S, Jayakumar ND, et al. Comparative estimation of sulfiredoxin levels between chronic periodontitis and healthy patients - A case-control study. *J Periodontol* 2018; 89: 1241–1248.
29. Kavarthapu A, Thamaraiselvan M. Assessing the variation in course and position of inferior alveolar nerve among south Indian population: A cone beam computed tomographic study.

Indian J Dent Res 2018; 29: 405–409.

30. Ramesh A, Ravi S, Kaarthikeyan G. Comprehensive rehabilitation using dental implants in generalized aggressive periodontitis. *J Indian SocPeriodontol* 2017; 21: 160–163.
31. Jain M, Nazar N. Comparative Evaluation of the Efficacy of Intraligamentary and Supraperiosteal Injections in the Extraction of Maxillary Teeth: A Randomized Controlled Clinical Trial. *J Contemp Dent Pract* 2018; 19: 1117–1121.
32. VijayashreePriyadharsini J. In silico validation of the non-antibiotic drugs acetaminophen and ibuprofen as antibacterial agents against red complex pathogens. *J Periodontol* 2019; 90: 1441–1448.
33. Amachawadi RG, Scott HM, Vinasco J, et al. Effects of in-feed copper, chlortetracycline, and tylosin on the prevalence of transferable copper resistance gene, *trcB*, among fecal enterococci of weaned piglets. *Foodborne Pathog Dis* 2015; 12: 670–678.
34. Solioz M. *Copper and Bacteria: Evolution, Homeostasis and Toxicity*. Springer, <https://play.google.com/store/books/details?id=5LZhDwAAQBAJ> (2018).
35. Elguindi J, Wagner J, Rensing C. Genes involved in copper resistance influence survival of *Pseudomonas aeruginosa* on copper surfaces. *J Appl Microbiol* 2009; 106: 1448–1455.
36. Grass G, Rensing C, Solioz M. Metallic copper as an antimicrobial surface. *Appl Environ Microbiol* 2011; 77: 1541–1547.
37. Santo CE, Quaranta D, Grass G. Antimicrobial metallic copper surfaces kill *Staphylococcus haemolyticus* via membrane damage. *MicrobiologyOpen* 2012; 1: 46–52.
38. Santo CE, Taudte N, Nies DH, et al. Contribution of Copper Ion Resistance to Survival of *Escherichia coli* on Metallic Copper Surfaces. *Applied and Environmental Microbiology* 2008; 74: 977–986.
39. Nandakumar R, Santo CE, Madayiputhiya N, et al. Quantitative proteomic profiling of the *Escherichia coli* response to metallic copper surfaces. *BioMetals* 2011; 24: 429–444.
40. Weber DJ, Otter JA, Rutala WA. Can Copper-Coated Surfaces Prevent Healthcare-Associated Infections? *Infection control and hospital epidemiology: the official journal of the Society of Hospital Epidemiologists of America* 2017; 38: 772–776.

UNDER PEER REVIEW